

## Total Protein (Single Reagent)

REF: 310 001 (2 x 100ml) 200 test  
 REF: 310 002 (4 x 100ml) 400 test  
 REF: 310 003 (8 x 100ml) 800 test  
 REF: 310 004 (2 x 500ml) 1000 test  
 REF: 310 005 (2 x 250ml) 500 test  
 REF: ZL-310 001 200 test

### Intended Use

Spectrum Total protein reagent is intended for the in-vitro quantitative, diagnostic determination of total protein in human serum or plasma on both automated and manual systems.

### Background

Plasma proteins are mainly synthesized in the liver and are involved in the maintenance of normal water distribution between tissues and blood, as well as acid-base balance. Due to some pathological conditions, both total protein level and the ratio of different fractions may change independently of one another. Hyperproteinemia may be detected during dehydration associated with diarrhea or vomiting. The total protein levels also increase in multiple myeloma. Hypoproteinemia may occur as a result of prolonged low protein diet and in some pathological conditions such as nephrotic syndrome, bleeding, sprue, and salt retention.

### Method

Colorimetric method (Biuret reagent).

### Assay Principle

In alkaline medium the copper reacts with the peptide bonds of proteins to form the characteristic pink to purple biuret complex. Sodium potassium tartarate prevents copper hydroxide precipitation, and potassium iodide prevents the autoreduction of copper.



The color intensity is directly proportional to the protein concentration. It is determined by measuring the increase in the absorbance at 546nm.

### Reagents

#### Standard Total protein (ST)

6.0 g/dL

#### Reagent (R)

Sodium hydroxide	750	mmol/L
Copper sulfate	12.0	mmol/L
Sodium potassium tartarate	40.9	mmol/L
Potassium iodide	19.8	mmol/L

(C)-Corrosive contains caustic materials.

**R34** Causes burns.

**S26-45** In case of contact with eyes, rinse immediately with plenty of water and seek medical advice in case of accident or if you feel unwell, seek medical advice immediately.

For further information, refer to the Total Protein reagent material safety data sheet.

### Precautions and Warnings

Do not ingest or inhale. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

### Reagent Preparation, Storage and Stability

Spectrum Total protein reagents are supplied ready-to-use and stable up to the expiry date labeled on the bottles. The reagents are stable at 15 – 25 °C. Only the standard needs to be kept refrigerated at (2 - 8°C). Once opened, the reagent is stable for 6 months and the standard is stable for 3 months at the specified temperature if contamination is avoided.

### SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Use by/Expiration Date
	For in-vitro diagnostic use		CAUTION. Consult instructions for use
	Batch Code/Lot number		Manufactured by
	Catalogue Number		(Xi) - Irritant
	Consult instructions for use		Temperature Limitation

### Deterioration

Do not use The total protein reagents if precipitate forms. Failure to recover control values within the assigned range may be an indication of reagent deterioration.

### Specimen Collection and Preservation

Use serum or plasma ( EDTA or heparin ) for the test . Usually plasma results are higher due to fibrinogen . The serum or plasma should be separated from the cells within 4 hours .  
 Stability : 1 day at 15 – 25 °C ; 4 weeks at 4 – 8 °C ;  
 1 year at -20 °C

### System Parameters

Wavelength	Hg 546 nm (530 – 570 nm)
Optical path	1 cm
Assay type	End-point
Direction	Increase
Sample : Reagent Ratio	1 : 50
e.g. : Reagent volume	1 ml
Sample volume	20 µl
Temperature	15 – 25 °C
Zero adjustment	Reagent blank
Incubation time	10 minutes at 15 – 25 °C
Reagent Blank Limits	Low 0.00 AU High 0.3 AU
Sensitivity	1.0 g/dL
Linearity	12 g/dL

### Procedure

	Blank	Standard	Sample
Reagent (R)	1.0 ml	1.0 ml	1.0 ml
Standard	-----	20 µl	-----
Sample	-----	-----	20 µl

Mix and Incubate for 10 minutes at room temp. Measure absorbance of specimen ( $A_{\text{specimen}}$ ) and standard ( $A_{\text{standard}}$ ) against reagent blank within 30 minutes.

### Calculation

$$\text{Serum protein conc. (g/dL)} = \frac{A_{\text{specimen}}}{A_{\text{standard}}} \times 6$$

**Note:** For turbid highly icteric sera, prepare a serum blank by adding 20 µl serum to 1 ml saline into a labeled test tube. Read absorbance of serum blank at 540 nm vs water and subtract serum blank absorbance from test absorbance before calculating results.

### Interfering Substances

#### Hemolysis

No interference up to hemoglobin level of 7.5 g/L.

#### Icterus

No significant interference up to a bilirubin level of 30 mg/dL.

#### Lipemia

No significant interference.

#### Drugs

Sera from patients receiving dextran may cause artificially high levels due to turbidity during color development. This positive bias can be minimized by centrifuging the reaction mixture before reading the absorbance.

## Performance Characteristics

### Precision

Within run (Repeatability)

	Level 1	Level 2
n	20	20
Mean (g/dL)	5.2	7.23
SD	0.12	0.15
CV%	2.31	2.07

Run to run (Reproducibility)

	Level 1	Level 2
n	20	20
Mean (g/dL)	5.7	7.32
SD	0.19	0.21
CV%	3.33	2.87

### Quality Control

Normal and abnormal commercial control serum of known concentrations should be analyzed with each run.

### Methods Comparison

A comparison between Spectrum Total protein reagent and a commercial reagent of the same methodology was performed on 200 human sera. A correlation of 0.978 was obtained.

### Sensitivity

When run as recommended, the minimum detection limit of this assay is 1.0 g/dL.

### Linearity

The reaction is linear up to total protein concentration of 12 g/dL. Specimens showing higher concentration should be diluted 1+1 using physiological saline and repeat the assay (result × 2).

### Expected Values

Adults		6.6 – 8.7 g/dL
Children	(> 1 year)	6.0 – 8.0 g/dL
	(< 1 year)	4.8 – 7.6 g/dL
Newborns	(< 4 weeks)	4.6 – 6.8 g/dL
Prematures		3.4 – 5.0 g/dL

### Analytical Range

1.0 – 12 g/dL.

**Spectrum Diagnostics does not interpret the results of a clinical laboratory procedure; interpretation of the results is considered the responsibility of qualified medical personnel. All indications of clinical significance are supported by literature references.**

## Waste Disposal

This product is made to be used in professional laboratories. Please consult local regulations for a correct waste disposal.

**S56:** dispose of this material and its container at hazardous or special waste collection point.

**S57:** use appropriate container to avoid environmental contamination.

**S61:** avoid release in environment. refer to special instructions/safety data sheets.

## References

1. Cannon DC, Olitzky I, Inkpen JA :Proteins. In:Clinical chemistry, principles and technics, 2 nd ed. RJ Henery, DC Cannon, JW Winkelman, editors, Harper & Row, New York, pp 407 – 421,1974.
2. Gornall AG, Bardawill CJ, David MM: Determination of serum protein by means of the biuret reagent. J Biol Chem 177:751, 1949 .
3. Kaplan A, Szalbo J :Clinical chemistry :Interpretation and techniques, 2 nd ed. A Kaplan, J Szabo, editors, 1983, p 157.
4. Schultze HE, Heremans JF:Molecular biology of human protein. Elsevier publishing company, Amsterdam, 1966.
5. Tietz NW : Fundamentals of Clinical Chemistry: 2 nd ed. NW Tietz, editor, 1994, pp692 .

## ORDERING INFORMATION

CATALOG NO.	QUANTITY
310 001	2 x 100 ml
310 002	4 x 100 ml
310 003	8 x 100 ml
310 004	2 x 500 ml
310 005	2 x 250 ml
REF: ZL-310 001	200 test



**Egyptian Co for Biotechnology - Spectrum Diagnostics (S.A.E)**

Obour city industrial area. block 20008 piece 19 A. Cairo. Egypt.

Tel: +202 4489 2248 - Fax: +202 4489 2247

www.spectrum-diagnostics.com

E-mail:info@spectrum-diagnostics.com



**MDSS GmbH**  
Schiffgraben 41  
30175 Hannover, Germany



IFUFCC38

Rev.(6), 6/6/2021