

Urea/BUN – single reagent

REF: 320 001	(2 x 20 ml)	40 test
REF: 320 002	(6 x 20 ml)	120 test
REF: 320 003	(4 x 50 ml)	200 test
REF: 320 004	(2 x 100 ml)	200 test
REF: ZL-320 001		40 test

Intended Use

Spectrum Ultimate Urea reagent is intended for the in-vitro quantitative, diagnostic determination of urea in human serum or urine on both automated and manual applications.

Background

Urea is the major product of protein nitrogen metabolism. It is synthesized by the urea cycle in the liver and excreted through the kidneys. The circulating levels of urea depend upon protein intake, protein catabolism and kidney function. Elevated urea levels can occur due to renal impairment or in some diseases such as diabetes, infection, congestive heart failure and during different liver diseases. Determination of blood urea nitrogen is the most widely used screening test for renal function together with serum creatinine.

Method

Urease-UV fixed rate (enzymatic method).

Assay Principle

The series of reactions involved in the assay are as follows :

1. Urea is hydrolyzed in the presence of water and urease to produce ammonia and carbon dioxide.



2. In the presence of glutamate dehydrogenase (GLDH) and reduced nicotinamide adenine dinucleotide (NADH), the ammonia combines with α -ketoglutarate (α -KG) to produce L-glutamate.



The rate decrease in the NADH concentration is directly proportional to the urea concentration in the specimen. It is determined by measuring the absorbance at 340 nm.

Reagents

Standard urea (ST)

BUN	50 mg/dL
Urea	107 mg/dL

Reagent

Tris Buffer (pH 8.5)	50 mmol/L
α -Ketoglutarate	10 mmol/L
GLDH	8.0 K U/L
Urease	5.0 K U/L
Sodium azide	8.0 mmol/L
NADH	>0.20 mmol/L
Sodium azide	8 mmol/L

The reagent also contains additives required to maintain NADH in its reduced form.

For further information, refer to the Urea/Bun reagent material safety data sheet.

Precautions and Warnings

Do not ingest or inhale. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

Reagent (R) contains sodium azide which may react with copper or lead plumbing.

SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Use by/Expiration Date
	For in-vitro diagnostic use		CAUTION. Consult instructions for use
	Batch Code/Lot number		Manufactured by
	Catalogue Number		(Xi) - Irritant
	Consult instructions for use		Temperature Limitation

Reagent Preparation , Storage and Stability

Spectrum Ultimate Urea reagent is supplied ready-to-use and stable up to the expiry date labeled on the bottles when stored refrigerated at 2-8 °C. Once opened, the reagent is stable for 2 months and the standard is stable for 3 months at the specified temperature.

Deterioration

Do not use Spectrum Ultimate Urea reagent if it is turbid or if the absorbance of the reagent is less than 0.9 at 340 nm. Failure to recover control values within the assigned range may be an indication of reagent deterioration.

Specimen Collection and Preservation

No special preparation of the patient is required. Use nonhemolyzed serum or plasma only. The only acceptable anticoagulants are heparin, EDTA and fluoride. Do not use ammonium heparin plasma.

Stability: 7 days at 15 –25 °C ; 7 days at 2 – 8 °C;
1 year at -20 °C

Urine samples are prediluted 1 : 50 with ammonium free water prior to assay.

Stability: 2 days at 15 – 25 °C ; 7 days at 2 – 8 °C;
1 month at -20 °C

System Parameters

Wavelength	340 nm
Optical path	1 cm
Assay type	Fixed Rate
Direction	Decrease
Sample : Reagent Ratio	1 : 100
e.g.: Reagent volume	1 ml
Sample volume	10 μ l
First read time	30 seconds
Delay time	60 seconds
Last read time	90 seconds
Temperature	37 °C
Zero adjustment	Against Air
Reagent Blank Limits	Low 0.9 AU High 2.0 AU
Sensitivity	0.9 mg/dL (0.15 mmol/L)
Linearity	200 mg/dL (33.2 mmol/L)

Procedure

	Standard	Specimen
Reagent (R)	1 ml	1 ml
Standard	10 μ l	-----
Specimen	-----	10 μ l

Mix and after 30 seconds read the absorbance A1 of the standard or specimen. Exactly 1 minute later, read the absorbance A2 of standard or specimen.

Calculation

$$\Delta A \text{ specimen} = A1 \text{ specimen} - A2 \text{ specimen}$$

$$\Delta A \text{ standard} = A1 \text{ standard} - A2 \text{ standard}$$

$$\text{Serum urea concentration (mg/dl)} = \frac{\Delta A_{\text{specimen}}}{\Delta A_{\text{standard}}} \times n$$

where $n = 107.0 \text{ mg/dL}$

Urine urea concentration is determined by multiplying the result by the dilution factor (50).

Quality Control

Normal and abnormal commercial control serum of known concentrations should be analyzed with each run.

Performance Characteristics

Precision

Within run (Repeatability)

	Level 1	Level 2
n	20	20
Mean (mg/dL)	45	150
SD	0.7	2.7
CV%	1.56	1.8

Run to run (Reproducibility)

	Level 1	Level 2
n	20	20
Mean (mg/dL)	47	153
SD	0.82	2.81
CV%	1.74	1.84

Methods Comparison

A comparison between Spectrum Ultimate Urea (UV) reagent and a commercial reagent of the same methodology was performed on 200 human sera. A correlation of 0.992 was obtained.

Sensitivity

When run as recommended, the minimum detection limit of the assay is 0.9 mg/dL.

Linearity

The reaction is linear up to a urea concentration of 200 mg/dL. Specimens showing higher concentration should be diluted 1+2 with physiological saline and repeat the assay (result x 3).

Interfering Substances

Haemolysis

Erythrocyte contamination doesn't elevate results. Haemolytic specimens may cause high absorbance flagging.

Icterus

No significant interference.

Lipemia

Lipemic specimens may cause high absorbance flagging. Diluted sample treatment may be recommended.

Anticoagulants

Ammonium heparin should not be used.

Others

Ammonium ions should be avoided since it may cause erroneously elevated results.

Expected Values

Urea (Serum)

Adults <65 years : 15-50 mg/dL (2.5-8.33 mmol/L)
Adults >65 years : < 70 mg/dL (<11.66 mmol/L)

BUN (Serum)

Adults <65 years : 7-23.5 mg/dL
Adults >65 years : 7-32.9 mg/dL
Children : 5-18 mg/dL

Urine (24) hours

Urea : 20 - 35 g/24hrs (330-580 mmol/24hrs)
BUN : 9.3-16.4 g/24hrs

Spectrum Diagnostics does not interpret the results of a clinical laboratory procedure; interpretation of the results is considered the responsibility of qualified medical personnel. All indications of clinical significance are supported by literature references.

Analytical Range

0.9 – 200 mg/dL (0.15 - 33.2 mmol/L).

Waste Disposal

This product is made to be used in professional laboratories. Please consult local regulations for a correct waste disposal.

S56: dispose of this material and its container at hazardous or special waste collection point.

S57: use appropriate container to avoid environmental contamination.

S61: avoid release in environment. refer to special instructions/safety data sheets.

References

1. Batton, C. J & Crouch, S.R: Anal. Chem. , 1977,49:464-469 .
2. Shephard MD, Mezzachi RD: Clin Biochem Revs, 4:61-7, 1983.
3. Tiffany TO, jansen JM, Burtis CA, Overtion JB, SCOTT CD. Enzymatic kinetic rate and end point analyses of substrate, by use of a gemsac fast analyzer. Clin Chem. 1972;18:829-840.
4. Tietz NW, Ed. Clinical guide to laboratory tests. 2ND. Philadelphia: WB Saunders;1990:566.

ORDERING INFORMATION

CATALOG NO.	QUANTITY
320 001	2 x 20 ml
320 002	6 x 20 ml
320 003	4 x 50 ml
320 004	2 x100 ml
ZL-320 001	40 test



Egyptian Co for Biotechnology - Spectrum Diagnostics (S.A.E)

Obour city industrial area. block 20008 piece 19 A. Cairo. Egypt.

Tel: +202 4489 2248 - Fax: +202 4489 2247

www.spectrum-diagnostics.com

E-mail: info@spectrum-diagnostics.com



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany



IFUFCC48

Rev.(5), 6/6/2021